

In Reference To: O-1320
Case No.: 45671 SDG(s): F6R29

**Contract Laboratory Program
REGIONAL/LABORATORY COMMUNICATION SYSTEM**

Resubmission Request

Laboratory Name:	EQI
Lab Contact:	Robert Zhu
Region:	6
Regional Contact:	Raymond Flores - EPA
ESAT Reviewer:	Wallace Doong - ESAT

In reference to data for the fraction(s):

TVOA/TVOA-SIM SVOA-SIM

Summary of Questions/Issues:

TVOA and TVOA-SIM

Sample F6R33: The 1,1,2-trichloroethane was reported at 9.0 ug/L in the undiluted TVOA analysis, but was not detected in the TVOA diluted analysis F6R33DL and TVOA-SIM analysis F6R33. Please review the data and make the appropriate resubmission.

SVOA-SIM

For Form 8A (pages 1162 and 1163), the IS3 should be phenanthrene-d10 and results for IS6 (1,4-dichlorobenzene-d4) were omitted for all samples. Please correct and resubmit these pages with proper pagination.

NOTE: Any laboratory resubmission should be submitted either as an addendum to the original CSF with a revised Form DC-2 or submitted as a new CSF with a new Form DC-2 except for replacement pages (SOM02.3, p. B-11, sec. 2.2.2). Custody seals are required only for regular mail shipments.

Please respond to the above items **within 5 business days** (SOM02.3, p. B-11, sec. 2.2.1) by e-mail to Flores.Raymond@epa.gov. If you have any questions, please contact Mr. Flores at 281-983-2139.

Distribution: (1) Lab Copy, (2) Region Copy, and (3) ESAT Copy